

NKS-436 ISBN 978-87-7893-526-7

Determination of important alpha emitters in nuclear samples – OptiMethod 2019 project report

Xiaolin Hou¹⁾ Maria Anderot²⁾ Lina Ekerljung, Susanna Wijk³⁾ Susanna Salminen-Paatero⁴⁾ Grzegorz Olszewski, Mats Eriksson, Patric Lindahl⁵⁾ Helene Öhlin, Hanna Åberg⁶⁾ Maria Davidsson, Erika Settervik⁷⁾ Filippa Bruzell, Tommy Suutari⁸⁾ Annika Tovedal, Anna Vesterlund⁹⁾ Satu Rautio, Satu Kangas¹⁰⁾ Stefan Allard¹¹⁾ Antti Hatakka. Petri Hovi 12) 1) Technical university of Denmark (DTU), Department of Environmental Engineering, Denmark 2) Swedish Nuclear Fuel and Waste Management Co, Clab (SKB), Sweden 3) Forsmarks Kraftgrupp AB, Sweden 4) University of Helsinki, Finland 5) Swedish Radiation Safety Authority (SSM), Sweden 6) Oskarshamn NPP, OKG Aktiebolag, Sweden 7) Ringhals AB, Sweden 8) Cyclife Sweden AB, Sweden 9) Swedish Defence Research Agency (FOI), Sweden 10) Loviisa NPP, Fortum, Power and Heat Oy, Finland 11) Chalmers University of Technology, Sweden. 12) Olkiluoto Nuclear Power Plant, Finland

May 2020

Abstract

This report presents the achievement of the NKS-B OptiMethod 2019 project which was conducted in 2019. An intercomparison was organized on determination of important alpha emitters in two real samples collected from Swedish nuclear power plants, one is nuclear fuel pool water and another is aerosol filter sample. Ten Nordic labs participated in this exercise and reported the analytical results of ²³⁸Pu, ^{239,240}Pu, ²⁴¹Am, ²⁴²Cm and ^{243,244}Cm and ²¹⁰Po. For the nuclear fuel pool water sample, the analytical results of ²³⁸Pu, ^{239, 240}Pu, ²⁴¹Am and ^{243, 244}Cm reported by all participating labs agree with each other very well, and the performance of all participating labs is much better than the intercomparison in 2018. Meanwhile the problems occurred in the intercomparison analysis in 2018, such as low chemical yields of Pu and Am/Cm, black residues in the prepared samples, and black and low quality of electrodeposition alpha source in some samples and in some labs were not observed in the analysis in 2019, all these indicate a significantly improved quality of the radiochemical analysis of these radionuclides in all partners' labs through this project. For the aerosol filter sample, the reported analytical results of alpha emitting radionuclides are lower than the detection limits of the methods in most of partners' labs, and a few reported data vary largely. This is mainly attributed to the very low concentrations of target radionuclides in the aerosol filter and relative small sample size used for this intercomparison analysis. The strategies for the analysis of alpha emitters in the samples with very high ²¹⁰Po content and the problem solution in the Nordic labs are also discussed in this report.

Key words: radioanalysis; nuclear power plant;pool water, aerosol filter, inter-comparison; plutonium; americium; curium

Determination of important alpha emitters in nuclear and samples –OptiMethod 2019 project report

Xiaolin Hou¹⁾, Maria Anderot²⁾, Lina Ekerljung³⁾ Susanna Salminen-Paatero⁴⁾, Grzegorz Olszewski⁵⁾ Helene Öhlin⁶⁾, Maria Davidsson⁷⁾, Filippa Bruzell⁸⁾, Annika Tovedal⁹⁾, Satu Rautio¹⁰⁾, Stefan Allard¹¹⁾, Antti Hatakka¹²⁾, Susanna Wijk³⁾, Mats Eriksson⁵⁾, Hanna Åberg⁶⁾, Erika Settervik⁷⁾, Tommy Suutari⁸⁾, Anna Vesterlund⁹⁾, Patric Lindahl⁵⁾, Satu Kangas¹⁰⁾, Petri Hovi¹²⁾

1) Technical university of Denmark, (DTU), Department of Environmental Engineering, Denmark

- 2) Swedish Nuclear Fuel and Waste Management Co, Clab (SKB), Sweden
- 3) Forsmarks Kraftgrupp AB, Sweden
- 4) University of Helsinki, Finland
- 5) Swedish Radiation Safety Authority (SSM), Sweden
- 6) Oskarshamn NPP, OKG Aktiebolag, Sweden
- 7) Ringhals AB, Sweden
- 8) Cyclife Sweden AB, Sweden
- 9) Swedish Defence Research Agency (FOI), Sweden
- 10) Loviisa NPP, Fortum, Power and Heat Oy, Finland
- 11) Chalmers University of Technology, Sweden.
- 12) Olkiluoto Nuclear Power Plant, Finland

Acknowledgement

NKS conveys its gratitude to all organizations and persons who by means of financial support or contributions in kind have made the work presented in this report possible.

Content

- 1. Introduction
- 2. Intercomparison exercise on determination of isotopes of Pu, Am and Cm in water samples
- 3. Problems and strategies on the analysis of alpha emitters
- 4. Summary and remarks

References

1. Introduction

Isotopes of plutonium (^{238, 239, 240}Pu), americium (²⁴¹Am) and curium (²⁴²Cm, ²⁴³Cm and ²⁴⁴Cm) are the most important anthropogenic radionuclides in radiation protection, environmental radioactivity, decommissioning of nuclear facilities and repository of nuclear waste, due to their relatively high production rates in the nuclear activities and high radiological toxicity of the alpha emitters.

Isotopes of Pu, Am and Cm are produced mainly in the nuclear fuel in the nuclear reactors and nuclear weapons tests. Some small amount of them might also be present in the reactor materials (through the reactions of impurity uranium with neutrons). Monitoring and reporting of these alpha emitters in atmospheric and liquid effluents from nuclear power plants are recommended by European Union (Euratom, 2004). This is also required by some authorities in Nordic countries in order to estimate the amount of these alpha emitters released to the environment and its impact to the publics and environment. Meanwhile, process water (reactor coolant and nuclear fuel pool water) and ion exchange resin are often analyzed in the NPPs to monitor the possible leakage of the nuclear fuel, to investigate the dispersion and deposition of alpha emitters in the nuclear reactor system, and to estimate the inventory of alpha emitters in the waste.

Fuel leakage has occurred during the operation of many nuclear power reactors, including those in the Nordic nuclear power plants such as Loviisa, Olkiluoto, Forsmark, Ringhals and Oskarshamn NPPs. It is essential to understand how the radionuclides were released from the fuel and to know the age of the leaked fuel, as well as to understand the behavior of the released radionuclides in the reactor system. Determination of the artificial alpha emitters is important for understanding of their behavior in the reactor water and their distribution in the reactor system, possible discharges to the environment and exposure to the operation staff and potentially to the public. These measurements can be also used to verify the simulation model currently used in the NPPs and to provide the important and critical information for decision-making. Therefore, the most Nordic NPPs have established radioanalytical laboratory for the determination of the alpha emitting radionuclides besides the beta and gamma emitters.

In the investigation of environmental radioactivity and radioecology, isotopes of Pu and Am in the environment and food are often determined to evaluate radiation exposure to the public and consequent radiation impact. Determination of alpha emitters is also an essential requirement in the nuclear and radiological emergency preparedness. In this case, a rapid analysis has to be implemented to meet the requirement of quick decision. For these purposes, sequential chemical separation and quick measurement are important. The Nordic authorities such as SSM, NRPA and STUK, and institutions such as DTU, University of Helsinki, FOI, etc. are heavily involved in these works.

Different methods have been developed and applied for determination of artificial alpha emitters in Nordic laboratories (Hou et al. 2016, 2019). In the Nordic nuclear industry laboratories such as all Nordic NPPs, SKB etc., two major methods are applied. One method is based on directly electrodeposition of metals on metal discs from the water sample and measurement by alpha spectroscopy. The major drawback with this method is that all alpha emitters with similar alpha energies (e.g. ²⁴¹ Am and ²³⁸ Pu, ²¹⁰Po and ²⁴³Am), cannot be discriminated, and the results are normally not radionuclide specific. For samples with high ²¹⁰Po, ²⁴³Am cannot be used as tracer for the determination of isotopes of Am and Cm. Another method is based on the chemical separation using Eichrom procedures, followed by electrodeposition and alpha spectrometry measurement. This method enables to separate each element and measure individual radionuclides, e.g. ²⁴¹ Am and ²³⁸ Pu. But it is high labor intensive and often associated with some problems, such as low separation efficiency, black residue occurring in the electrodeposition, white residue formed in the separated solution, causing bad resolution in the prepared alpha source, unsuitable chemical yield tracers, no precise adjustment of the electrodeposition solution. The major problem is lack of the validation of the analytical method and verification of the analytical results for different samples, because suitable standard reference materials and inter-laboratory comparisons for relevant sample matrix are not available. Consequently, the NPPs labs have a strong requirement for validation and improvement of their analytical methods for determination of alpha emitters in their routine analysis. In addition, the behavior of alpha emitters released during the leakage in the reactor is still not well understood.

The OptiMethod project was initiated in 2018, aims to summarize and improve the analytical methods used in Nordic labs for the determination of alpha emitting radionuclides. The analytical capacity and methods used in Nordic labs for determination of alpha emitters have been reviewed and summarized (Hou et al. 2019). 11 Nordic labs in Denmark, Finland and Sweden have capacities to determine individual alpha emitters including ²³⁸Pu, ^{239, 240}Pu, ²⁴¹Am, ²⁴²Cm and ^{243, 244}Cm in nuclear (reactor and spent fuel pool water) and/or environmental samples. Alpha spectrometry is used in most of labs for the measurement, sequential separation using different combination of ion exchange and extraction chromatography, e.g. UTEVA-TRU, TEVA-UTEVA-TRU, Dowex-TRU and TEVA-DGA columns are used for separation of Pu, Am and Cm from matrices and each other. These methods were summarized in the project report (Hou et al. 2019). An inter-comparison exercise for analysis of a spiked water and a real reactor water was organized in 2018, 11 Nordic labs participated in this inter-comparison exercise and reported the analytical results of isotopes of plutonium, americium and curium. The results reported by the most labs agree with each other with only 1-2 outlying data for each radionuclide (Hou et al. 2019). Some problems were recognized in this inter-comparison exercise:

1) The direct electrodeposition followed by alpha spectrometry method was used in one Nordic lab for the determination of alpha emitters. In this case, it is not possible to report the individual activity concentrations of ²³⁸Pu and ²⁴¹Am because of the similar energies of alpha particles emitted from these two radionuclides. For the analysis of samples with complicated matrices (e.g. discharge water, environmental samples), this method might be not usable.

- 2) Variable and low recovery of Am and Cm (down to 18-50%) in the analytical procedure was observed in some labs for some samples. Meanwhile, low recoveries of Pu (10-20%) in the analytical procedure in some labs were also reported. Besides the chemical separation procedure, electrodeposition step might be another reason.
- 3) It was reported that the prepared source using electrodeposition is not always good, sometimes a black source was obtained, causing a bad energy resolution in alpha spectra in some samples. Besides the electrodeposition condition, the purity of the separated sample solution might be another reason, which might be also a reason causing low recovery of the radionuclides measured in the alpha sources. A further investigation and optimization to overcome this problem is needed.
- 4) A black residue was found in the separated Am-Cm fraction after evaporation in some lab using UTEVA-TRU separation procedure, it was not clear how the residue was formed and if it contains Am and Cm. Organic residues eluted from the UTEVA and TRU resin might be the possible source of this black residue. A further investigation is needed to clarify it, and to assess its effect in the separation of Am and Cm.
- 5) Different separation procedures are used in the Nordic labs for the separation of alpha emitters, all of them work well for the analysis of reactor water samples with simple matrices. However, the suitability of these procedures for analysis of samples with complicated matrices, such as filter paper and discharge water from nuclear power plants, environmental samples such as soil, sediment, seawater and fresh water is not known. One challenge might be the removal of ²¹⁰Po interference in the analysis of filter and environmental samples. The concentration of ²¹⁰Po in filter sample might be a few orders of magnitude higher compared to the measured alpha emitters, the similar energy of alpha particles with ²¹⁰Po (5.30 MeV), namely ²⁴³Am (5.28 MeV) which was used as a yield tracer of Am and Cm isotopes makes the analytical results questionable if ²¹⁰Po was not well eliminated. Another challenge will be the sufficient removal of the large amount of matrix elements in the environmental samples. A further improvement and investigation for optimization of the analytical method for isotopes of U, Pu, Am and Cm is needed.

The OptiMethod2019 project is a continuation of the work in 2018, it aims to solve the problem recognized in the last stage of the project, to improve the analytical quality of the Nordic laboratories for radiochemical analysis of artificial alpha emitters in NPP, waste and environmental samples by organizing the second inter-comparison exercises, in order to evaluate and optimize method used in participating labs.

2. Inter-comparison exercise on determination of isotopes of Pu, Am and Cm in nuclear fuel pool water and aerosol filter of a nuclear power plants

The second inter-comparison exercise was organized in this project in 2019 for the determination of isotopes of Pu, Am and Cm in a nuclear fuel pool water sample and an aerosol filter samples collected from one Nordic NPP. Ten Nordic labs participated in the inter-comparison exercise. The analytical results of this inter-comparison exercise are presented and discussed below.

2.1 Samples

One nuclear fuel pool water and one aerosol filter sample collected from one NPP were prepared for this exercise.

- (1) Nuclear fuel pool water was collected from one of Swedish Nuclear fuel storage pools in Oct. 2018, and provided by SKB. This sample was pure water without addition of boric acid or other chemicals. The sample was acidified to about pH2 using HNO₃ immediately after sampling. 2 liters of water sample in a plastic container were sent to partner laboratories for inter-comparison exercise. This sample contains some activation and fission product radionuclides, with total gamma activity of about 100 Bq/L. The concentrations of ^{239,240}Pu, ²⁴¹Am and ^{243,244}Cm were estimated to be 0.1-10 mBq/L of each radionuclide, and 1-50 mBq/L for ²³⁸Pu.
- (2) Aerosol was collected on 0.45µm membrane filters of 47 mm in diameter at different parts of the ventilation systems in a Swedish nuclear power pant, more than 30 filters were collected in 2018, which were used for this inter-comparison exercise. To avoid the problem of inhomogeneity of radionuclides among filters and in each filter, all collected filters were combined to one sample and cut into small pieces. The filters were leached using 1M H₂SO₄ solution for 3 times by heated on a hotplate. All leachates were combined and filtered through a 0.45 micrometer filter, in total 1500 ml leachate was obtained. After well mixed, the leachate was divided into aliquots of 100 ml (equivalent to 2 filters), sealed into plastic bottle, and delivered to each partners lab for inter-comparison exercise. The concentrations of 1-10 mBq/100 ml for each isotope of Pu, Am and Cm and 0.1-2.0 Bq/100ml for ²¹⁰Po were estimated.

Sample	Sample description	Matrix	Preparation/	Alpha radionuclides to be	
code			collection date	determined and major	
				radionuclides, activation and small	
				fission products	
Pool water	Nuclear fuel pool	2.0 L, acidified to	Oct. 2018	²³⁸ Pu, ²³⁹ Pu, ²⁴⁰ Pu, ²⁴¹ Pu, ²⁴² Pu,	
	water from SKB	pH2 with HNO ₃		²⁴¹ Am, ^{243,244} Cm	
Filter	Leachate of aerosol	100 ml leachate of	March - Nov.	²³⁸ Pu, ²³⁹ Pu, ²⁴⁰ Pu, ²⁴¹ Pu, ²⁴¹ Am,	
	filter collected from	1M H ₂ SO ₄ ,	2018	²⁴³ Am, ²⁴² Cm, ²⁴³ Cm ²⁴⁴ Cm, other	
	Forsmark NPP	filtered through		activation and fission product	
		0.45 μm filter		radionuclides, relative high ²¹⁰ Po	
				level.	

Table 1 Two inter-comparison water samples for determination of alpha emitters

3.2 Analytical methods applied for the determination of alpha emitters by the participants

In this inter-comparison exercise, the measurement of alpha emitters was implemented by alpha spectrometry in all participating partner labs. The alpha sources were mainly prepared by electrodeposition on stainless steel disc in most of labs, but a few labs also used micro-precipitation method by forming co-precipitation of actinide fluoride with LnF₃, the precipitates was transferred to a small filter paper by filtration and used for alpha measurement. For the chemical separation, 4 different procedures were used by the 10 participating labs. Among them, 6 labs used UTEVA-TRU method (Fig. 1), 4 of them used electrodeposition for alpha source preparation and two labs used micro-precipitation (NdF_3) for source preparation. One lab directly electrodeposited actinide on stainless steel disc for alpha measurement of the radionuclides without chemical separation. The sample solution was simply evaporated to dryness and the residue was dissolved into H₂SO₄ solution and then adjust to pH2.1-2.4 for electrodeposition of all actinide on to disc. Other 3 labs used different chemical separation procedures, including TEVA-UTEVA-TRU-Ln method (Fig. 2), TEVA-DGA method (Fig. 3), and TEVA-TRU-TEVA method. The detailed procedures of these methods have been reported elsewhere (Hou et al. 2019). ²⁴²Pu and ²⁴³Am purchased from different organizations (e.g. NIST, NPL, Eckert & Ziegler) were used as yield tracers for plutonium isotopes and isotopes of Am and Cm, respectively in the chemical separation in all labs who did the chemical separation. 0.01-0.08 Bq of ²⁴²Pu and ²⁴³Am were spiked to sample solution before chemical separation. In the direct measurement method (no chemical separation was applied), ²³³U (0.1 Bq) was spiked as yield tracer for estimation of the recoveries of all alpha radionuclides during sample preparation and electrodeposition processes, and for correcting the activity

of all other radionuclides in alpha spectrometry measurement. Table 2 summarizes the methods used in this comparison by participating labs.

Lab code	Separation method	Source preparation	Tracer(s)	Measurement methods
1	TEVA-DGA	Electrodeposition	²⁴² Pu, ²⁴³ Am	α-spectrometry
4	TEVA-TRU-TEVA	Electrodeposition	²⁴² Pu, ²⁴³ Am	α-spectrometry
5, 7, 10, 11	UTEVA-TRU	Electrodeposition	²⁴² Pu, ²⁴³ Am	α-spectrometry
6	TEVA-UTEVA-	Electrodeposition	²⁴² Pu, ²⁴³ Am	a-spectrometry
	TRU-Ln resin			
8	no	Electrodeposition	²³³ U	α-spectrometry
12, 9	UTEVA-TRU	Micro-coprecipitation	²⁴² Pu, ²⁴³ Am	α-spectrometry

Table 2 Analytical methods applied in partners' lab for inter-comparison exercise



Fig. 1 Diagram of an analytical procedure for determination of Pu, Am and Cm with a combination of UTEVA and TRU columns used in 6 partner labs.



Fig.2 Diagram of an analytical procedure for determination of Pu, Am and Cm with a combination of TEVA, UTEVA, TRU and Ln columns used in a Nordic lab.



Fig. 3 Diagram of chromatographic separation procedure for sequential separation of Pu, Np, Am and Cm and measurement of their isotopes.

2.3 Analytical results of inter-comparison exercise

(1) Nuclear fuel pool water

Overall 10 labs participated in the inter-comparison analysis of the nuclear fuel pool water sample, among them, 10 labs reported analytical results of the isotopes of Am and Cm, 9 labs reported the results of ²³⁸Pu and ^{239,240}Pu. The analytical results with uncertainties (k=2) of each lab are presented in Fig. 4. All results were decay corrected to 1st Jan. 2019. The values of average (red line) and 2SD (yellow lines) of all reported data are also presented in the figures.

The reported ²³⁸Pu concentrations in this sample range from 16.9 to 21.3 mBq/kg with an average and 1SD of 19.4 ± 2.0 mBq/kg. The variation of the analytical results among all participating labs is relative small with a relative standard deviation of only 10%. These results are much better than the intercomparison in 2018 although the concentration of ²³⁸Pu in this sample is 2 times lower than that in the reactor water used in the inter-comparison in 2018.

The reported concentrations of ^{239,240}Pu in the pool water range from 0.52 to 1.60 mBq/L with an average and 1SD of 1.10±0.39 mBq/kg. A relative bigger variation among participating labs with a relative standard deviation of 36% was obtained. This is mainly attributed to the low concentration of ^{239,240}Pu in this sample, more than 10 times lower than that of ²³⁸Pu. However, there is no exceptional

value which are far from the most of data in this inter-comparison for ^{239,240}Pu, indicating the performance on the determination of ^{239,240}Pu also better than that in 2018, especially in consideration of more than 2 orders of magnitude lower ^{239,240}Pu in this sample compared to the reactor water samples (527±59 mBq/kg in average and 1SD) used in the inter-comparison in 2018.



Fig. 4 Inter-comparison results of nuclear fuel pool water sample for ²³⁸Pu, ^{239,240}Pu, ²⁴¹Am and ²⁴⁴Cm in Nordic labs.

The reported concentrations of ²⁴¹Am in the pool water sample range from 4.06 to 5.30 mBq/kg with an average and 1SD of 4.71±0.46 Bq/kg. The variation of the analytical results among all participating

labs is relative small with a relative standard deviation of only 10%. These results are also much better than in the inter-comparison in 2018 although the concentration of 241 Am in this sample is 40 times lower than that in the reactor water (254±39 mBq/kg) used in the inter-comparison in 2018.

The reported concentrations of ^{243,244}Cm in the pool water range from 1.40 to 2.87 mBq/kg with an average and 1SD of 2.27±0.44 mBq/kg. The variation of the results among all participating labs is 20%, which is comparable with the results of ^{243,244}Cm of the inter-comparison in 2018 (1500-2400 mBq/kg), although the concentration of ^{243,244}Cm in the pool water in this inter-comparison is nearly 3 orders of magnitude lower than that in the reactor water used in the inter-comparison in 2018.

In general, the analytical results of all participating labs for ²³⁸Pu, ^{239,240}Pu, ²⁴¹Am and ^{243, 244}Cm in the nuclear fuel pool water agree very well. Although the concentrations of these radionuclides in the pool water samples are 2 times to 3 orders of magnitude lower than those in the reactor coolant water used in the inter-comparison in 2018, the performance of all participating labs is better than that in 2018, indicating a significantly improved performance of the radiochemical analysis in the Nordic labs. This should be attributed to the activities of organized in this project, which help the participants to solve the problems and to improve the analytical method and the operation skills in the radiochemical analysis.

For ²⁴²Cm, the most of the reported results are below the detection limit of the methods used by the partners (0.5 mBq/kg), only a few labs reported values (e.g. 0.06 mBq/kg) above the detection limits of their methods. Therefore it is difficult to make a comparison for this radionuclide in this sample, and the results are therefore not presented in the figure. ²⁴²Cm is a relative short-lived (162 days) radionuclide, which is produced mainly through the beta decay of the shortlived ²⁴²Am (16 h) and internal conversion followed by beta decay of relative long-lived ^{242m}Am (141 years) (Fig. 5). The nuclear fuel pool water used in this inter-comparison was collected in Oct. 2018, and the sample was analyzed in each participating lab in July - Sept. 2019, i.e. more than 240 days after collection. The spent fuels stored in the pool might has already been taken out from the reactor for a few years. Therefore, the contribution from ²⁴²Am should be negligible, and the dominant source of ²⁴²Cm should be ^{242m}Am. The decay correction for ²⁴²Cm (about half year) might not have a significant contribution to the variation of the results.



Fig. 5 Production routes of isotopes of Am and Cm in the nuclear reactor (Holm et al. 2002) The undetectable level of ²⁴²Cm in the nuclear fuel pool water indicates that its production and reservation in the nuclear fuel pool is very small. The concentrations of alpha emitting isotopes of plutonium, americium and curium in the fuel pool water analyzed in this inter-comparison are significantly lower than those in the reactor water analyzed in the inter-comparison in 2018, indicating that no leakage of the fuel elements might be occurred in this nuclear fuel pool, the trace amounts of these alpha emitting actinides might originate from the impurities of uranium presented on the surface of the fuel elements, which was exposed to the neutrons during the utilization of the fuel in the power reactor.

(2) Aerosol filter from nuclear power plant

The aerosol filters collected from different parts of ventilation system in a Swedish nuclear power plant was used in the inter-comparison in this work. To avoid the homogeneity problem, all 30 filter samples were cut and leached using H₂SO₄, and the leachate was filtered and divided into 100 ml aliquots to be distributed to the partners' labs. The analytical results of the radionuclides were reported in mBq/kg of leachate.

Table 2 shows the analytical results of ²³⁸Pu, ^{239,240}Pu, ²⁴¹Am, ²⁴²Cm and ^{243,244}Cm. It can be seen that the reported results are lower than the detection limits in the most of the labs. This is mainly attributed to the very low concentrations of these radionuclides in the air released through the ventilation system from the nuclear power plant. Meanwhile, the sample size used for inter-comparison exercise was too small to be capable of measuring the very low concentrations of alpha emitting radionuclides. The relative high detection limits of these radionuclides in the analysis of this sample compared to those for the nuclear fuel pool water sample is mainly attributed to the small size of filter leachate of only 30 -100 g compared to 1000 g for the pool water sample. This results in a much less amount of radionuclides on the electrodeposited disc. For a 50 g filter leachate sample, with 90% chemical yield, 25% counting efficiency of alpha spectrometry, and a detection limit of 0.1 counts/h of the alpha spectrometry, the detection limit of the method will be 2.5 mBq/kg. Fig. 6 shows the alpha spectra of plutonium fraction and Am/Cm fraction with ²⁴²Pu and ²⁴³Am as yield tracers, respectively. It can be seen that no ²³⁸Pu and ^{239,240}Pu peaks in the plutonium fraction and no ²⁴¹Am, ²⁴²Cm, ^{243,244}Cm peaks in the Am/Cm fraction were measureable, and only the peaks of ²⁴²Pu and ²⁴³Am (which were spiked to the sample as yield tracers) in the spectra were visible, even when the sources were measured for 7 days.

Lab	²³⁸ Pu, m	Bq/kg	^{239,240} Pu, mBq/kg		²⁴¹ Am, mBq/kg		²⁴² Cm, mBq/kg		^{243,244} Cm, mBq/kg	
code										
	Value	Uncert	Value	Uncert	Value	Uncert	Value	Uncert	Value	Uncert
1	<5.2		<5.2		<3.0		<2.5		<2.8	
4	<2		<2		0.85	0.17	0.98	0.18	0.58	0.13
5	29.40	8.30	<6.7		9.20	0.75	32.9	10.6	<3.3	
6										
7	<1.9		<1.9		2.70	0.71	1.90	1.30	<0.71	
8	45.20	3.62	4.70	1.03	with ²³⁸ Pu	1		9.30	1.488	
10	<25.6		<8.8		12.60	4.80	<6.9		<2.6	
11	<14		<14		<21		<15		<11	
12	<12.8		<8.5		<6.9		<13.0		<3	

Table 2 Results of inter-comparison analysis for isotopes of Pu, Am and Cm in filter leachate

*The data presented as "< xx " are the detection limits of the methods used in the partner's labs. The analytical uncertainties presented are extended uncertainty with k=2.



Fig. 6 Alpha spectra of plutonium fraction and Am and Cm fraction determined in the nuclear fuel pool water samples for inter-comparison 2019.

The very low concentrations of isotopes of Pu, Am and Cm in the aerosol filter should be attributed to the refractory features of the actinides, which are stable in the circumstances in the nuclear reactor, not volatile, and which doesn't easily form aerosol even presented in the reactor water. These results also indicate that the atmospheric releases is not a significant pathway for the isotopes of Pu, Am and Cm.

A few labs reported the analyzed values for some radionuclides, but showed a very large variation (Table 2). This might be attributed to the small size of the sample used for the analysis and the efficiency for decontamination of the interference radionuclides might be not sufficiently high. It was observed that ²¹⁰Po concentration in the aerosol sample is quite high. Three labs have analyzed the ²¹⁰Po concentration in the filter leachate, the reported concentrations of ²¹⁰Po ranged in 2530-3370 mBq/kg with an average and 1SD of 2870±440 mBq/kg. This is 2-3 orders of magnitude higher than those of alpha isotopes of Pu, Am and Cm. The chemical properties of ²¹⁰Po are very complicated, and ²¹⁰Po might enter to the fractions of plutonium and Am/Cm fraction, and cause a high interference, especially for the determination of isotopes of Am and Cm, because the alpha energy of ²¹⁰Po (5.30 MeV) is close to that of ²⁴³Am (5.23-5.28 MeV), and the alpha spectrometry could not discriminate these two radionuclides. Since ²⁴³Am was employed as yield tracer for Am and Cm, the contamination of ²¹⁰Po in the Am/Cm fraction will cause an overestimated values of ²⁴¹Am, ²⁴²Cm and ^{234, 244}Cm.

4. Problems and strategies in the analysis of alpha emitters

1) Challenge and strategies in the analysis of filter samples from a nuclear power plant

The selection of the aerosol filter samples from the nuclear power reactor for the inter-comparison exercise in this work was based on the following considerations: 1) atmospheric releases of radionuclides in the nuclear power plants is an important issue in view of radiation protection, and the monitoring of the radioactivity level in the effluents including atmospheric releases is required by the authorities of radiation protection; 2) The Nordic labs have demonstrated a good capacity in the analysis of alpha radionuclides of Pu, Am and Cm in the inter-comparison exercise in 2018 by analysis of spiked water and the reactor coolant water, which allows to plan a more challenging work for determination of low level alpha emitting isotopes of Pu, Am and Cm in the second stage of the project in 2019; 3) Some labs of Nordic nuclear industry had started and planned for the analysis of filter samples for alpha emitting isotopes of Pu, Am, and Cm. It was known that the concentrations of alpha emitting radionuclides in the aerosol filters are very low, and the activity concentration of ²¹⁰Po in this type of filter is a few orders of magnitude higher than the isotopes of actinides. Therefore, an investigation was implemented for better elimination of the interferences of ²¹⁰Po in the determination of these radionuclides, especially in the Am/Cm fraction. Meanwhile, the analytical

procedures used for the routine analysis were modified to eliminate polonium in the chemical separation procedure.

In the chemical procedure with UTEVA-TRU chromatographic columns (Fig. 1) which was used by 6 partner's labs, polonium can be removed in the TRU column separation step. Po^{4+} , Pu^{3+} , Am^{3+} and Cm^{3+} do not adsorb on the UTEVA column at 3-4 M HNO₃ media. The prepared sample solution in 3M HNO₃ is loaded to a UTEVA column, where Po^{4+} , as well as Pu^{3+} , Am^{3+} and Cm^{3+} pass through the UTEVA column and load to the TRU column. Po^{2+} could not be adsorbed in the TRU resin in HNO₃ solution, but is highly adsorbed in HCl media (Fig. 7). On the contrary, Pu^{3+} , Am^{3+} and Cm^{3+} have a relatively high adsorption on TRU resin in high concentrations of HNO₃ (>1M) media (Fig.7). After loading the sample solution in 3M HNO₃ media to the TRU column, the TRU column was rinsed with high concentration of HNO₃ (e.g. 4-8 M) with addition of NaNO₂ to on-column oxidizing Pu^{3+} to Pu^{4+} to be strongly adsorbed on TRU resin, then extensively rinsing the TRU column with 8 M HNO₃ to highly eliminate Po^{4+} remaining on the column. Afterwards, Am/Cm and Pu were sequentially eluted using 4M HCl and complexing reagent (NH₄)₂C₂O₄, respectively. Since the Am/Cm was eluted with 4M HCl and high adsorption of Po⁴⁺ on TRU resin occurs at HCl media, a high decontamination factor of Po⁴⁺ in the Cm/Am fraction is expected by this modification.



Fig. 7 Distribution factor of Po⁴⁺ Pu³⁺, Am³⁺ and Cm³⁺ on TRU resin (Maxwell et al. 2013; Triskem 2020).

In the TEVA-DGA procedure used in one partner lab (Fig. 3), the procedure was modified to remove Po⁴⁺ in both TEVA and DGA chromatographic separation steps. Po⁴⁺ has a relative high distribution factors on TEVA resin in low concentrations of HNO₃ and HCl media, especially in HCl medium (Fig. 8). However, the distribution factors of Po⁴⁺ on the TEVA decrease with the increased concentrations of the HCl and HNO₃, 1-2 orders of magnitude lower distribution factor of Po⁴⁺ on the TEVA resin could be obtained in 8 M HNO₃ media. Based on this feature, the prepared sample solution in 4M HNO₃ media is loaded onto the TEVA column, followed by rinsing with 4M HNO₃. In this case Pu⁴⁺ and Po⁴⁺ are highly retained on the TEVA column, while Am/Cm pass through the TEVA column and enter to the effluent and rinse solution for further purification. Additional step rinsing using 8M HNO₃ was added in the modified procedure after the loading and rinsing TEVA column (Fig. 9), in order to get a better removal of Po⁴⁺ from the Pu fraction.

In the DGA resin, Po⁴⁺ has a high adsorption in HCl media, but relatively low and increasing distribution factor on the DGA resin with increasing concentrations of HNO₃. The distribution factor of Po⁴⁺ is about 2 at 0.1M HNO₃ media (Fig. 10). The Am/Cm fraction loaded in a DGA column was further rinsed with 0.2 M HNO₃ to remove any remaining Po⁴⁺ in the effluent from the TEVA column before eluting Am/Cm with HCl solution, to ensure a high decontamination factor for ²¹⁰Po in the separated Am/Cm solution, and eliminate the interference of ²¹⁰Po to the measurement of ²⁴³Am (tracer).



Fig. 8 Variation of k' values of Po⁴⁺ and other ions with the concentration of HNO₃ and HCl in TEVA resin (Triskem 2020, Khaing & Thakur, 2017)



Fig. 9 Modified procedure for the separation of Pu, Am and Cm from aerosol filter samples with high ²¹⁰Po content



Fig. 10 Fig. 8 Variation of k' values of Po⁴⁺ and other ions with the concentration of HNO₃ and HCl in DGA resin (Triskem, 2020; Maxwell et al. 2019)

2) Improvement of the preparation of alpha sources for the measurement of Am and Cm isotopes.

In the inter-comparison exercise in 2018, some labs reported that a black residue was observed in the separated Am-Cm fraction after addition of H_2SO_4 and evaporation to dryness using UTEVA-TRU separation procedure, It was proposed that the black residue might be organic residues from the TRU or/and UTEVA resin. A modified procedure was used in this lab by evaporation of the Am/Cm fraction to dryness without addition of H_2SO_4 . The formed residue was dissolved with 0.05M H_2SO_4 solution for electrodeposition. By modification, no black residue was observed anymore, and the chemical yield of Am/Cm monitored by ²⁴³Am was not significantly changed. This might confirm that the black residue is the organic substance from the resin, which could be burned to blank carbon by concentrated H_2SO_4 at high temperature during evaporation step, after HCl and H2O were removed from the eluate. However, it should be mentioned that the complete removal of HCl and HNO₃ is necessary before electrodeposition, otherwise it might interfere with the electrodeposition of Am/Cm on the stainless steel.

It was also reported in the inter-comparison in 2018 that a black disc was observed Θ in some lab during electrodeposition and consequently a bad energy resolution in alpha spectra in these samples occurred. This was attributed to the possible impurities in the separated Am/Cm fraction, including organic substances which might be burned during electrodeposition at high electric current (1.0-1.5 A). In the inter-comparison in 2019, no such as black alpha source was observed, meanwhile, the chemical yields of Am/Cm and Pu in the corresponding fraction were also sufficiently high (> 60%), no recovery of less than 20% was observed. This demonstrated that the investigation and discussion through the implementation of the project practically helped the partners for improvement of their analytical procedure and experimental skills. With the effort of the participating labs in the operation of the chemical separation procedure, a better removal of the impurities and improved chemical yield of Pu and Am/Cm were ensured.

5. Summary and Remarks

This project is a continuation of the work in 2018. Based on the achievement and knowledge obtained during the operation of the project in 2018, the second run inter-comparison analysis of two real samples was organized, ten Nordic labs participated in this inter-comparison exercise and reported the results of isotopes of plutonium, americium and curium. The problems and strategies identified in 2018 were investigated for improvement of the performance of the radiochemical analysis in the Nordic labs for reliable determination of these radionuclides. The major achievements are summarized below:

 All 12 partner labs of this project in Denmark, Finland and Sweden have established the capability for the determination of alpha emitters ²³⁸Pu, ^{239, 240}Pu, ²⁴¹Am, ²⁴²Cm and ^{243, 244}Cm in nuclear (reactor and spent fuel pool water) samples. In most of the labs, alpha spectrometry is used for the measurement of these alpha emitting radionuclides. ICP-MS is only used in a few labs for measurement of ²³⁹Pu, ²⁴⁰Pu and isotopes of uranium (²³⁴U, ²³⁵U, ²³⁸U), mainly in the labs of universities and institutes.

- 2) All reported results of ²³⁸Pu, ^{239, 240}Pu, ²⁴¹Am and ^{243, 244}Cm by the participating labs in the intercomparison exercise of the nuclear fuel pool water samples agree very well, although the activity concentrations of these radionuclides are significantly lower by a factor of up to 3 orders of magnitude than that in the reactor coolant water used in the inter-comparison in 2018. This indicates a significantly improved performance in the Nordic labs in the radiochemical analysis.
- 3) The problems of black residue and black electrodeposition sources occurred in the inter-comparison analysis in some partners' labs in 2018 did not occur in the analysis in 2019. Besides the modification of the analytical procedure, the experimental skills in the separation were also significantly improved. This was also confirmed by the high chemical yields of Pu and Am/Cm in all participating labs.
- 4) An aerosol filter sample collected from the ventilation system in a Swedish nuclear power plant was also used for the inter-comparison exercise. Unfortunately, the reported concentrations of ²³⁸Pu, ^{239, 240}Pu, ²⁴¹Am, ²⁴²Cm and ^{243, 244}Cm are lower than the detection limits of the method used in most of the participating labs. This is mainly attributed to the very low concentrations of these radionuclides in the air in the nuclear power plant, and the relative small size of the sample used for the analysis.
- 5) Although the concentrations of the interesting radionuclides in the filter sample were not measurable, however, the main challenge of very high ²¹⁰Po content compared to the interesting alpha emitter in this type of samples was well recognized in the partners' labs, and the analytical procedures were modified to improve the elimination of the ²¹⁰Po interference to the measurement of isotopes of Am and Cm.
- 6) Besides water and filter samples from nuclear power plants, various environmental samples (soil, sediment, seawater, vegetation, and animal tissues) and waste samples from decommissioning of nuclear facilities are also being analyzed in many Nordic labs. Therefore validation, verification and improvement of the analytical methods for these sample types in the Nordic labs are also highly needed. A continuation of this project is expected to increase the reliability of the present applied analytical method for alpha emitting radionuclides in these types of samples in the Nordic labs. A new NKS project of RAD-MERDE was proposed and launched in 2020. In this project, the improved analytical method achieved in the OptiMethod project for sequentially analyzing Pu, Am, and Cm in NPP water samples will be expanded to facilitate the analysis of different solid environmental and decommissioning samples.

References:

- Euratom (2004). Commission Recommendation of 18 December 2003 on standardisation on radioactive airborn and liquid discharges into the environment from nuclear power reactors and reprocessing plants in normal operation, C(2003)4832, 2004/2/Euratom.
- Holm E., Roos P., Aarkrog A., Mitchell P., Vintro L.L. (2002). Curium isotopes in Chernobyl fallout. J. Radioanal. Nucl. Chem. 252:211-214.
- Hou X.L., Allard S., Hatakka A., Eriksson M., Salminen-Paatero S., Bruzell F., Ekerljung L., Andersson A., Öhlin H., Anderot M., Tovedal A., Rautio S., Hansson N., Hovi P., Olszewski G., Suutari T., Davidsson M., Åberg H., Kangas S. (2019) Simultaneous Determination of Isotopes of Pu, Am and Cm in Reactor Water Samples –Report of OptiMethod 2018 project, NKS-415, ISBN 978-87-7893-504-5, Nordic Nuclear Safety Research, 2019. http://www.nks.org/scripts/getdocument.php?file=111010214696185
- HouX.L., Olsson M., Togneri L., Englund S., Vaaramaa K., Askeljung C., Gottfridsson O., Hirvonen H., Öhlin H., Forsström M., Anders F., Lampén M., Hatakka A. (2016). Present status and perspective of radiochemical analysis of radionuclides in Nordic countries. J. Radioanal. Nucl., Chem. 309:1283-1319.
- Khaing H., Thakur P. (2017). Rapid sequential separation method for 210Po and actinides in air filter samples. J. Radioanal. Nucl. Chem., 314:1383-1392
- Maxwell S.L., Culligan B.L., Hutchison J.B., Utsey R.C., McAlister D.R. (2013). Rapid determination of 210Po in water samples. J. Radioanal. Nucl. Chem., 298:1977-1989.
- Maxwell S.L., McAlister D.R., Sudowe R. (2019). Rapid method to determine polonium-210 in urban matrices. Appl. Radiat. Iso. 148:270-276.
- Triskem (2020). Extraction chromatography, Technical Documents. Triskem International, 2020. https://www.triskem-international.com/scripts/files/5addcf96423962.97324869/technical_doc_allproducts_web-0.pdf

Bibliographic Data Sheet

Title	Determination of important alpha emitters in nuclear and environmental samples –OptiMethod 2019 project report
Author(s)	Xiaolin Hou ¹⁾ , Maria Anderot ²⁾ , Lina Ekerljung ³⁾ Susanna Salminen-Paatero ⁴⁾ , Grzegorz Olszewski ⁵⁾ , Helene Öhlin ⁶⁾ , Maria Davidsson ⁷⁾ , Filippa Bruzell ⁸⁾ , Annika Tovedal ⁹⁾ , Satu Rautio ¹⁰⁾ , Stefan Allard ¹¹⁾ , Antti Hatakka ¹²⁾ , Susanna Wijk ³⁾ , Mats Eriksson ⁵⁾ , Hanna Åberg ⁶⁾ , Erika Settervik ⁷⁾ , Tommy Suutari ⁸⁾ , Anna Vesterlund ⁹⁾ , Patric Lindahl ⁵⁾ , Satu Kangas ¹⁰⁾ , Petri Hovi ¹²⁾
Affiliation(s)	1)Technical university of Denmark, (DTU), Denmark; 2) Swedish Nuclear Fuel and Waste Management Co, Clab (SKB), Sweden; 3)Forsmarks Kraftgrupp AB, Sweden; 4)University of Helsinki, Finland; 5)Swedish Radiation Safety Authority (SSM), Sweden; 6) Oskarshamn NPP, OKG Aktiebolag, Sweden; 7)Ringhals AB, Sweden; 8)Cyclife Sweden AB, Sweden; 9)Swedish Defence Research Agency (FOI), Sweden; 10)Loviisa NPP, Fortum, Power and Heat Oy, Finland; 11)Chalmers University of Technology, Sweden; 12)Olkiluoto Nuclear Power Plant, Finland
ISBN	978-87-7893-526-7
Date	May 2020
Project	NKS-B / OptiMethod
No. of pages	23
No. of tables	2
No. of illustrations	10
No. of references	8
Abstract max. 2000 characters	This report presents the achievement of the NKS-B OptiMethod 2019 project which was conducted in 2019. An intercomparison was organized on determination of important alpha emitters in two real samples collected from Swedish nuclear power plants, one is nuclear fuel pool water and another is aerosol filter sample. 10 Nordic labs participated in this exercise and reported the analytical results of isotopes of plutonium, americium, curium and ²¹⁰ Po. For the nuclear fuel pool water sample, the analytical results of ²³⁸ Pu, ^{239, 240} Pu, ²⁴¹ Am and ^{243, 244} Cm reported by all participating labs agree with each other very well, and the performance of all participating labs is much better than in the intercomparison in 2018. Meanwhile the problems occurred in the intercomparison analysis of 2018, such as low chemical yield of Pu and Am/Cm, black residues in the prepared samples, and black and low quality of

	electrodeposition alpha source were not observed in the analysis in 2019. All
	these indicate a significantly improved quality of the radiochemical analysis of
	these radionuclides through this project. While, for the aerosol filter sample,
	the reported analytical results of alpha emitting radionuclides are lower than
	the detection limits of the method used in the partners' lab, and a few reported
	data also vary largely, this is mainly attributed to the very low concentrations
	of target radionuclides in the aerosol filter and relatively small sample size for
	the analysis. The strategies for the analysis of alpha emitters in sample with
	very high 210Po content and the problem solution in the Nordic labs are also
	discussed in this report.
Key words	Alpha emitter; Radioanalysis; reactor water; Inter-comparison; Plutonium;
	Americium; Curium

Available on request from the NKS Secretariat, P.O.Box 49, DK-4000 Roskilde, Denmark. Phone (+45) 4677 4041, e-mail nks@nks.org, www.nks.org